

Nitrite CHEMets[®] Kit

K-7004/R-7002: 0 - 2.5 ppm N

Safety Information

Read SDS (available at www.chemetrics.com) before performing this test procedure. Wear safety glasses and protective gloves.

Test Procedure

1. Fill the sample cup to the 25 mL mark with the sample to be tested (fig. 1).
2. Place the CHEMet ampoule, tip first, into the sample cup. Snap the tip. The ampoule will fill leaving a bubble for mixing (fig. 2).
3. To mix the ampoule, invert it several times, allowing the bubble to travel from end to end.
4. Dry the ampoule and wait **10 minutes** for color development.
5. Obtain a test result by placing the ampoule between the color standards until the best color match is found (fig. 3).

NOTE: To convert to ppm nitrite (NO₂), multiply test result by 3.3.




Figure 1




Figure 2




Figure 3

Test Method

The Nitrite CHEMets[®] test kit employs the azo dye formation method.^{2,3} In an acidic solution, nitrite diazotizes with a primary aromatic amine and then couples with another organic molecule to produce a highly colored azo dye. The resulting pink-orange color is proportional to the nitrite concentration in the sample.

1. CHEMets is a registered trademark of CHEMetrics, Inc. U.S. Patent No. 3,634,038
2. APHA Standard Methods, 22nd ed., Method 4500-NO₂⁻ B -2000
3. EPA Methods for Chemical Analysis of Water and Wastes, Method 354.1 (1983).

Visit www.chemetrics.com to view product demonstration videos.
Always follow the test procedure above to perform a test.



Simplicity in Water Analysis

www.chemetrics.com

4295 Catlett Road, Midland, VA 22728 U.S.A.

Phone: (800) 356-3072; Fax: (540) 788-4856

E-Mail: orders@chemetrics.com

Mar. 18, Rev. 8